

Technical Compliance Statement EMC Test Report

Ref. File No.: C1M2310262

For the following information

Product : Bed Mover

Model Name : GZ10SL

Brand : Fallshaw

Applicant : Fallshaw Wheels & Castors

Manufacturer : Fallshaw Wheels & Castors

Standards :

EN 60601-1-2:2015 +A1:2021 (IEC 60601-1-2:2014 +AMD:2020)

Emission: CISPR 11:2015 +A1:2016 +A2:2019, Class B Immunity: IEC 61000-4-2:2008, IEC 61000-4-3:2020

IEC 61000-4-8:2009, IEC 61000-4-39:2017

We hereby certify that the above product has been tested by us with the listed standards and found in compliance with the Regulation (EU) 2017/745 concerning medical device. The test data and results are included in the test report **EM-E240017**.

Signature

Alex Deng/Deputy Manager

Date: 2024. 01. 09

Test Laboratory:

Audix Technology Corporation, EMC Department

TAF Accreditation No.: 1724 Web Site: www.audixtech.com

The statement is based on a single evaluation of one sample of the above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab logo.



Technical Compliance Statement EMC Test Report

Ref. File No.: C1M2310262

For the following information

Product : Bed Mover

Model Name : GZ10SL

Brand : Fallshaw

Applicant : Fallshaw Wheels & Castors

Manufacturer : Fallshaw Wheels & Castors

Standards :

EN 60601-1-2:2015 +A1:2021 (IEC 60601-1-2:2014 +AMD:2020)

Emission: CISPR 11:2015 +A1:2016 +A2:2019, Class B Immunity: IEC 61000-4-2:2008, IEC 61000-4-3:2020

IEC 61000-4-8:2009, IEC 61000-4-39:2017

We hereby certify that the above product has been tested by us with the listed standards and found in compliance with the Regulation (EU) 2017/745 concerning medical device. The test data and results are included in the test report **EM-E240018**.

Signature

Alex Deng/Deputy Manager

Date: 2024, 01, 11

Test Laboratory:

Audix Technology Corporation, EMC Department

TAF Accreditation No.: 1724 Web Site: www.audixtech.com

The statement is based on a single evaluation of one sample of the above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab logo.